

B. Biocompatibility

Old Item No.	Standard	Change	Replacement Item No.
7	ASTM F719–81(2002)e1, Standard Practice for Testing Biomaterials in Rabbits for Primary Skin Irritation.	Withdrawn and replaced with newer version.	68
30	ASTM F720–81(2002)e1, Standard Practice for Testing Guinea Pigs for Contact Allergens: Guinea Pig Maximization Test.	Withdrawn and replaced with newer version.	69
32	ASTM F750–87(2002)e1, Standard Practice for Evaluating Material Extracts by Systemic Injection in the Mouse.	Withdrawn and replaced with newer version.	70

C. Cardiovascular/Neurology

Old Item No.	Standard	Change	Replacement Item No.
46	ASTM F2079–02 Standard Test Methods for Measuring Recoil of Balloon-Expandable Stents.	Recognize newer year date version.	49

D. Dental/ENT

Old Item No.	Standard	Change	Replacement Item No.
48	ANSI/ADA Specification No. 16:1989, Dental Impression Paste Zinc Oxide—Eugenol Type.	Correction in title (dash between oxide and eugenol).	
64	ISO 3107:1988, Dental Zinc Oxide/Eugenol Cements and Zinc Oxide Non-Eugenol Cements.	Correction in title (slash between oxide and eugenol).	
66	ISO 4049:1988, Dentistry-Resin—Based Filling Materials.	Correction in year date (1988 instead of 1998).	
86	ANSI/ADA Specification No. 38:2000, Metal-Ceramic Systems.	Correction in title (change to systems).	

E. General

Old Item No.	Standard	Change	Replacement Item No.
6	IEC 60601–1–2, (First Edition, 1993–04), Medical Electrical Equipment—Part 1: General Requirements for Safety; Electromagnetic Compatibility—Requirements and Tests.	Re-recognize	6
28	IEC 60601–1–2, (Second Edition), Medical Electrical Equipment—Part 1: General Requirements for Safety; Electromagnetic Compatibility - Requirements and Tests.	Extension of time period for the transition statement.	28

H. General Hospital/ General Plastic Surgery

Old Item No.	Standard	Change	Replacement Item No.
81	ASTM E1061	Title correction	81