

Old Item No.	Standard	Change	Replacement Item No.
5	ASTM F138–00, Standard Specification for Wrought 18 Chromium–14 Nickel–2.5 Molybdenum Stainless Steel Bar and Wire for Surgical Implants (UNS S31673).	Cardiovascular contact person. Clarification to extent of recognition with regard to biocompatibility requirements.	
6	ASTM F139–00, Standard Specification for Wrought 18 Chromium–14 Nickel–2.5 Molybdenum Stainless Steel Sheet and Strip for Surgical Implants (UNS S31673).	Clarification to extent of recognition with regard to biocompatibility requirements.	
7	ASTM F560–98, Standard Specification for Unalloyed Tantalum for Surgical Implant Applications (UNS R05200, UNS R05400).	Cardiovascular contact person. Clarification to extent of recognition with regard to biocompatibility requirements.	
8	ASTM F562–02, Standard Specification for Wrought 35 Cobalt–35 Nickel–20 Chromium–10 Molybdenum Alloy for Surgical Implant Applications (UNS R30035).	Withdrawn and replaced with newer version. Cardiovascular contact person. Clarification to extent of recognition with regard to biocompatibility requirements.	45
9	ASTM F563–00, Standard Specification for Wrought Cobalt–20 Nickel–20 Chromium–3.5 Molybdenum–3.5 Tungsten–5 Iron Alloy for Surgical Implant Applications (UNS R30563).	Cardiovascular contact person. Clarification to extent of recognition with regard to biocompatibility requirements.	
10	ASTM F603–00, Standard Specification for High-Purity Dense Aluminum Oxide for Surgical Implant Application.	Clarification to extent of recognition with regard to biocompatibility requirements.	
11	ASTM F620–00, Standard Specification for Alpha Plus Beta Titanium Alloy Forgings for Surgical Implants.	Clarification to extent of recognition with regard to biocompatibility requirements.	
12	ASTM F621–02, Standard Specification for Stainless Steel Forgings for Surgical Implants.	Withdrawn and replaced with newer version. Clarification to extent of recognition with regard to biocompatibility requirements.	46
13	ASTM F648–00, Standard Specification for Ultra-High-Molecular-Weight Polyethylene Powder and Fabricated Form for Surgical Implants.	Clarification to extent of recognition with regard to biocompatibility requirements.	
14	ASTM F688–00, Standard Specification for Wrought Cobalt–35 Nickel–20 Chromium–10 Molybdenum Alloy Plate, Sheet, and Foil for Surgical Implants (UNS R30035).	Clarification to extent of recognition with regard to biocompatibility requirements.	
15	ASTM F745–00, Standard Specification for 18 Chromium–12.5 Nickel–2.5 Molybdenum Stainless Steel for Cast and Solution-Annealed Surgical Implant Applications.	Clarification to extent of recognition with regard to biocompatibility requirements.	
17	ASTM F799–02, Standard Specification for Cobalt–28 Chromium–6 Molybdenum Alloy Forgings for Surgical Implants (UNS R31537, R31538, R31539).	Withdrawn and replaced with newer version. Clarification to extent of recognition with regard to biocompatibility requirements.	47
18	ASTM F899–02, Standard Specification for Stainless Steel for Surgical Instruments.	Withdrawn and replaced with newer version. Clarification to extent of recognition with regard to biocompatibility requirements.	48