

and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.

(e) Components of reagent kits of different lot numbers must not be interchanged unless otherwise specified by the manufacturer.

§ 493.1253 Standard: Establishment and verification of performance specifications.

(a) *Applicability.* Laboratories are not required to verify or establish performance specifications for any test system used by the laboratory before April 24, 2003.

(b)(1) *Verification of performance specifications.* Each laboratory that introduces an unmodified, FDA-cleared or approved test system must do the following before reporting patient test results:

(i) Demonstrate that it can obtain performance specifications comparable to those established by the manufacturer for the following performance characteristics:

(A) Accuracy.

(B) Precision.

(C) Reportable range of test results for the test system.

(ii) Verify that the manufacturer's reference intervals (normal values) are appropriate for the laboratory's patient population.

(2) *Establishment of performance specifications.* Each laboratory that modifies an FDA-cleared or approved test system, or introduces a test system not subject to FDA clearance or approval (including methods developed in-house and standardized methods such as text book procedures, Gram stain, or potassium hydroxide preparations), or uses a test system in which performance specifications are not provided by the manufacturer must, before reporting patient test results, establish for each test system the performance specifications for the following performance characteristics, as applicable:

(i) Accuracy.

(ii) Precision.

(iii) Analytical sensitivity.

(iv) Analytical specificity to include interfering substances.

(v) Reportable range of test results for the test system.

(vi) Reference intervals (normal values).

(vii) Any other performance characteristic required for test performance.

(3) *Determination of calibration and control procedures.* The laboratory must determine the test system's calibration procedures and control procedures based upon the performance

specifications verified or established under paragraph (b)(1) or (b)(2) of this section.

(c) *Documentation.* The laboratory must document all activities specified in this section.

§ 493.1254 Standard: Maintenance and function checks.

(a) *Unmodified manufacturer's equipment, instruments, or test systems.* The laboratory must perform and document the following:

(1) Maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.

(2) Function checks as defined by the manufacturer and with at least the frequency specified by the manufacturer. Function checks must be within the manufacturer's established limits before patient testing is conducted.

(b) *Equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer.* The laboratory must do the following:

(1)(i) Establish a maintenance protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting.

(ii) Perform and document the maintenance activities specified in paragraph (b)(1)(i) of this section.

(2)(i) Define a function check protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting.

(ii) Perform and document the function checks, including background or baseline checks, specified in paragraph (b)(2)(i) of this section. Function checks must be within the laboratory's established limits before patient testing is conducted.

§ 493.1255 Standard: Calibration and calibration verification procedures.

Calibration and calibration verification procedures are required to substantiate the continued accuracy of the test system throughout the laboratory's reportable range of test results for the test system. Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following:

(a) Perform and document calibration procedures—

(1) Following the manufacturer's test system instructions, using calibration materials provided or specified, and

with at least the frequency recommended by the manufacturer;

(2) Using the criteria verified or established by the laboratory as specified in § 493.1253(b)(3)—

(i) Using calibration materials appropriate for the test system and, if possible, traceable to a reference method or reference material of known value; and

(ii) Including the number, type, and concentration of calibration materials, as well as acceptable limits for and the frequency of calibration; and

(3) Whenever calibration verification fails to meet the laboratory's acceptable limits for calibration verification.

(b) Perform and document calibration verification procedures—

(1) Following the manufacturer's calibration verification instructions;

(2) Using the criteria verified or established by the laboratory under § 493.1253(b)(3)—

(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and

(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and

(3) At least once every 6 months and whenever any of the following occur:

(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes.

(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance.

(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem.

(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.

§ 493.1256 Standard: Control procedures.

(a) For each test system, the laboratory is responsible for having control procedures that monitor the accuracy and precision of the complete analytical process.

(b) The laboratory must establish the number, type, and frequency of testing control materials using, if applicable, the performance specifications verified