

each 24-hour period and the number of hours spent examining slides in the 24-hour period irrespective of the site or laboratory.

(4) Records are available to document the workload limit for each individual.

(e) *Slide examination and reporting.* The laboratory must establish and follow written policies and procedures that ensure the following:

(1) A technical supervisor confirms each gynecologic slide preparation interpreted to exhibit reactive or reparative changes or any of the following epithelial cell abnormalities:

(i) Squamous cell.

(A) Atypical squamous cells of undetermined significance (ASC-US) or cannot exclude HSIL (ASC-H).

(B) LSIL-Human papillomavirus (HPV)/mild dysplasia/cervical intraepithelial neoplasia 1 (CIN 1).

(C) HSIL-moderate and severe dysplasia, carcinoma in situ (CIS)/CIN 2 and CIN 3 or with features suspicious for invasion.

(D) Squamous cell carcinoma.

(ii) Glandular cell.

(A) Atypical cells not otherwise specified (NOS) or specified in comments (endocervical, endometrial, or glandular).

(B) Atypical cells favor neoplastic (endocervical or glandular).

(C) Endocervical adenocarcinoma in situ.

(D) Adenocarcinoma endocervical, adenocarcinoma endometrial, adenocarcinoma extrauterine, and adenocarcinoma NOS.

(iii) Other malignant neoplasms.

(2) The report of gynecologic slide preparations with conditions specified in paragraph (e)(1) of this section must be signed to reflect the technical supervisory review or, if a computer report is generated with signature, it must reflect an electronic signature authorized by the technical supervisor who performed the review.

(3) All nongynecologic preparations are reviewed by a technical supervisor. The report must be signed to reflect technical supervisory review or, if a computer report is generated with signature, it must reflect an electronic signature authorized by the technical supervisor who performed the review.

(4) Unsatisfactory specimens or slide preparations are identified and reported as unsatisfactory.

(5) The report contains narrative descriptive nomenclature for all results.

(6) Corrected reports issued by the laboratory indicate the basis for correction.

(f) *Record and slide retention.* (1) The laboratory must retain all records and slide preparations as specified in § 493.1105.

(2) Slides may be loaned to proficiency testing programs in lieu of maintaining them for the required time period, provided the laboratory receives written acknowledgment of the receipt of slides by the proficiency testing program and maintains the acknowledgment to document the loan of these slides.

(3) Documentation of slides loaned or referred for purposes other than proficiency testing must be maintained.

(4) All slides must be retrievable upon request.

(g) *Automated and semi-automated screening devices.* When performing evaluations using automated and semi-automated screening devices, the laboratory must follow manufacturer's instructions for preanalytic, analytic, and postanalytic phases of testing, as applicable, and meet the applicable requirements of this subpart K.

(h) The laboratory must document all control procedures performed, as specified in this section.

#### **§ 493.1276 Standard: Clinical cytogenetics.**

(a) The laboratory must have policies and procedures for ensuring accurate and reliable patient specimen identification during the process of accessioning, cell preparation, photographing or other image reproduction technique, photographic printing, and reporting and storage of results, karyotypes, and photographs.

(b) The laboratory must have records that document the following:

(1) The media used, reactions observed, number of cells counted, number of cells karyotyped, number of chromosomes counted for each metaphase spread, and the quality of the banding.

(2) The resolution is appropriate for the type of tissue or specimen and the type of study required based on the clinical information provided to the laboratory.

(3) An adequate number of karyotypes are prepared for each patient.

(c) Determination of sex must be performed by full chromosome analysis.

(d) The laboratory report must include a summary and interpretation of the observations, number of cells counted and analyzed, and use the International System of Cytogenetic Nomenclature.

(e) The laboratory must document all control procedures performed, as specified in this section.

#### **§ 493.1278 Standard: Histocompatibility.**

(a) *General.* The laboratory must meet the following requirements:

(1) An audible alarm system must be used to monitor the storage temperature

of specimens (donor and recipient) and reagents. The laboratory must have an emergency plan for alternate storage.

(2) All patient specimens must be easily retrievable.

(3) Reagent typing sera inventory prepared in-house must indicate source, bleeding date and identification number, reagent specificity, and volume remaining.

(4) If the laboratory uses immunologic reagents (for example, antibodies, antibody-coated particles, or complement) to facilitate or enhance the isolation of lymphocytes, or lymphocyte subsets, the efficacy of the methods must be monitored with appropriate quality control procedures.

(5) Participate in at least one national or regional cell exchange program, if available, or develop an exchange system with another laboratory in order to validate interlaboratory reproducibility.

(b) *HLA typing.* The laboratory must do the following:

(1) Use a technique(s) that is established to optimally define, as applicable, HLA Class I and II specificities.

(2) HLA type all potential transplant recipients at a level appropriate to support clinical transplant protocol and donor selection.

(3) HLA type cells from organ donors referred to the laboratory.

(4) Use HLA antigen terminology that conforms to the latest report of the World Health Organization (W.H.O.) Committee on Nomenclature. Potential new antigens not yet approved by this committee must have a designation that cannot be confused with W.H.O. terminology.

(5) Have available and follow written criteria for the following:

(i) The preparation of cells or cellular extracts (for example, solubilized antigens and nucleic acids), as applicable to the HLA typing technique(s) performed.

(ii) Selecting typing reagents, whether prepared in-house or commercially.

(iii) Ensuring that reagents used for typing are adequate to define all HLA-A, B and DR specificities that are officially recognized by the most recent W.H.O. Committee on Nomenclature and for which reagents are readily available.

(iv) The assignment of HLA antigens.

(v) When antigen redefinition and retyping are required.

(6) Check each HLA typing by testing, at a minimum the following:

(i) A positive control material.

(ii) A negative control material in which, if applicable to the technique performed, cell viability at the end of